Cureus

Review began 08/21/2023 Review ended 09/04/2023 Published 09/06/2023

#### © Copyright 2023

Kaneda et al. This is an open access article distributed under the terms of the Creative Commons Attribution License CC-BY 4.0., which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

# Pancreas Ligation Device for Distal Pancreatectomy: An Ex Vivo Follow-Up Porcine Study

Yuji Kaneda <sup>1, 2</sup>, Yuki Kimura <sup>1</sup>, Akira Saito <sup>1</sup>, Ryusuke Ae <sup>3</sup>, Hiroshi Kawahira <sup>1, 2</sup>, Naohiro Sata <sup>1</sup>

 Department of Surgery, Division of Gastroenterological, General and Transplant Surgery, Jichi Medical University, Shimotsuke, JPN 2. Medical Simulation Center, Jichi Medical University, Shimotsuke, JPN 3. Division of Public Health, Center for Community Medicine, Jichi Medical University, Shimotsuke, JPN

Corresponding author: Yuji Kaneda, ykaneda819@jichi.ac.jp

# Abstract

#### Introduction

Postoperative pancreatic fistula (POPF) is a critical complication occurring with a high incidence after distal pancreatectomy. To minimize the risk of POPF, we developed an innovative pancreas ligation device capable of closing the pancreatic stump without causing traumatic injury to the pancreatic duct and artery. We conducted an *ex vivo* follow-up study to compare the pressure resistance of the pancreas ligation device with that of a regular linear stapler.

#### Materials and methods

The pancreases were excised from 20 pigs and divided into two groups: ligation group (n = 10) and stapler group (n = 10). Distal pancreatectomy was performed, and the pancreatic stump was closed using either a pancreas ligation device or a regular linear stapler. The main pancreatic duct was cannulated with a 4-French catheter connected to a cannula and syringe filled with contrast medium. Using fluoroscopy detection, pressure resistance was defined as the maximum pressure without leakage from the pancreatic stump.

#### Results

No significant differences were found between the two groups regarding sex, age, body weight, or pancreatic thickness. In the ligation group, no leakage was observed at the stump in any pancreas. However, in the stapler group, six of 10 pancreases showed leakage at the staple line or into the parenchyma. Pressure resistance was significantly higher in the ligation group than in the stapler group (median: 42.8 vs. 34.3 mmHg, P = 0.023).

#### Conclusions

These findings suggest the effectiveness of a pancreas ligation device in reducing the incidence of POPF after distal pancreatectomy. Our ligation device is expected to be a useful alternative to a linear stapler for pancreatic stump closure.

Categories: Gastroenterology, General Surgery, Oncology

**Keywords:** bioabsorbable, innovation, pressure resistance, porcine study, pancreatic necrosis, pancreatic stump, complication, postoperative pancreatic fistula, distal pancreatectomy, pancreas ligation device

# Introduction

Postoperative pancreatic fistula (POPF) is a critical postoperative complication of distal pancreatectomy. POPF can cause fatal secondary complications, such as intra-abdominal hemorrhage, abscess, and sepsis [1-4], resulting in prolonged hospitalization [1,3,5] and increased medical costs [6,7]. Furthermore, POPF is associated with adjuvant chemotherapy failure [8,9], leading to a worsened prognosis in cases of pancreatic cancer [10,11].

Although several new techniques and devices have been introduced for distal pancreatectomy, the incidence of clinically relevant POPF [12] remains high at 15-22% [13-16]. Recently, reinforced staplers have been used for pancreatic stump closure; however, multicenter randomized clinical trials have not demonstrated any advantage of reinforced staplers over regular linear staplers for clinically relevant POPF prevention [17,18]. For these reasons, the optimal closure methods for pancreatic stump to avoid POPF complications after distal pancreatectomy remain controversial [19].

We previously proposed that the pancreas ligation device can potentially prevent POPF after distal

#### How to cite this article

Kaneda Y, Kimura Y, Saito A, et al. (September 06, 2023) Pancreas Ligation Device for Distal Pancreatectomy: An Ex Vivo Follow-Up Porcine Study. Cureus 15(9): e44771. DOI 10.7759/cureus.44771

pancreatectomy due to its atraumatic ligation approach, with the hypothesis that conventional closure methods for pancreatic stump (e.g., hand-sewn closure, regular linear stapler, and reinforced stapler) cause direct injury to the pancreatic parenchyma and pancreatic duct by staples or needles, thereby increasing the risk of POPF development [20].

To minimize the risk of POPF, we developed an innovative pancreas ligation device capable of closing the pancreatic stump without causing traumatic injury to the pancreatic duct and artery [20]. During the process of improving the device, we modified it from a band type [20] to a front-to-back clamping type, aiming to equalize the ligature pressure and transmit it to the center of the ligature.

To evaluate the effectiveness of the modified pancreas ligation device in reducing the risk of POPF, we conducted an *ex vivo* study comparing the pressure resistance of the main pancreatic duct (MPD) after distal pancreatectomy with a regular linear stapler. This was a follow-up study to verify the findings of the initial pilot *in vivo* study [20].

# **Materials And Methods**

## **Pancreas ligation device**

We developed an innovative pancreas ligation device for distal pancreatectomy with technical support from Teijin Medical Technologies Co., Ltd. (Osaka, Japan), made from a bioabsorbable polymer with four specific characteristics (Fig. *1A*) [20].



## FIGURE 1: Pancreas ligation device.

(A) The ligation device can ligate the pancreas from the front and back and has an adjustable locking system. (B) Image of the ligation device after distal pancreatectomy. The device ligates the pancreatic stump atraumatically.

Image credits: Yuji Kaneda

First, the device ligates the pancreas from the front and back without using any directly traumatic components, such as staples and needles (Fig. *1A*, *1B*). Second, an adjustable locking system controls the ligation force to maintain the arterial blood flow to the pancreatic stump (Fig. *1A*). Third, the grooves prevent the device from sliding or displacing the ligature site. Finally, the device is made of a bioabsorbable polymer, leaving behind no foreign bodies. We conducted this *ex vivo* study using a prototype of the device created from non-absorbable plastic with a three-dimensional printer.

## Ex vivo pancreas and distal pancreatectomy

This study was approved by the Animal Experiment Committee of our institution (approval no. 18035-02), and all animals were managed according to the ethical regulations for animal studies. Twenty pigs (10 female and 10 castrated male Landrace pigs) were purchased from Sanesu Breeding Co., Ltd. (Funabashi, Japan). The pancreases were individually excised under general anesthesia and thereafter were wrapped, stored at -25 °C, and finally were divided into two groups: ligation group (n = 10) and stapler group (n = 10).

Distal pancreatectomy was performed with defrosted pancreases, and the pancreatic stump was closed using a pancreas ligation device and regular linear stapler in the ligation and stapler groups, respectively. In the ligation group, we ligated the pancreas at 1 cm to the left of the confluence of the splenic vein and divided the pancreas with Metzenbaum scissors at 0.5 cm from the ligation line (Fig. *2A*, *2B*, *3A*, *3B*). The ligation force of the device was to the locking limit (Fig. *1A*, *3B*).



### FIGURE 2: Schematic illustration of porcine distal pancreatectomy.

(A) Ligation and division line of the pancreas. The ligation and division lines are set at 1 cm and 1.5 cm to the left of the confluence of the splenic vein. Pancreases in the stapler group were divided across the division line, (B) while those in the ligation group were first ligated across the ligation line and then divided across the division line.

MPD, main pancreatic duct

Image credits: Yuji Kaneda



#### FIGURE 3: Pancreatic stump closure with the pancreas ligation device.

(A) The pancreas is divided with Metzenbaum scissors. (B) The pancreatic stump is ligated with the pancreas ligation device.

In the stapler group, we divided the pancreas with a linear stapler (Endo GIA <sup>TM</sup> Black Reload with Tri-Staple<sup>TM</sup> Technology 60 mm; Covidien Japan, Inc., Japan) at 1.5 cm to the left of the splenic vein confluence (Fig. 2A, 4A, 4B).



### FIGURE 4: Pancreatic stump closure with a linear stapler.

(A) The pancreas is divided with a linear stapler. (B) The pancreatic stump is closed with staples.

### Pressure resistance test

The MPD pressure after distal pancreatectomy was compared between the ligation and stapler groups. The MPD was cannulated with a 4-French catheter placed through the duodenal lobe of the pancreas, which corresponds to the head of the pancreas in humans (Fig. 5).



# FIGURE 5: Schematic illustration of the pressure resistance test.

The MPD is cannulated through the duodenal lobe of the pancreas. The catheter is connected to the cannula and syringe by a three-way stopcock. The pancreas is cut along the cut line. Leakage from the pancreatic stump is detected by X-ray fluoroscopy. MPD, main pancreatic duct. Image credits: Yuji Kaneda.

The catheter was then connected to a cannula marked from 5 to 50 cm (at 5-cm intervals, 10 steps) and a syringe filled with 30% contrast medium. The cannula was filled with a contrast medium, and the MPD pressure was increased from 5 to 50 cm (at 5-cm intervals, 10 steps). This operation increased the pressure from 4.3 to 42.8 mmHg in 4.3-mmHg increments. After opening the stopcock, the MPD pressure was maintained for five seconds by holding the stopcock at each level. Contrast leakage from the pancreatic stump or staple line was detected using fluoroscopy. We defined the MPD pressure resistance as the maximum pressure without leakage.

#### **Statistical analyses**

The numerical variables are presented as the median (minimum to maximum). The Mann-Whitney U test was performed to compare the MPD pressure resistance values. The significance threshold was set at P < 0.05. All analyses were performed using the IBM SPSS Statistics software program for Windows, version 29 (released 2022, IBM Corp., Armonk, NY, USA).

# **Results**

No significant differences were found between the ligation and stapler groups in the sex (castrated male/female ratio: 6/4 vs. 4/6), age (94 (79 to 118) vs. 88 (76 to 110) days old), body weight (34.0 (27.8 to 40.2) vs. 33.8 (30.0 to 40.5) kg), or thickness of the pancreas (8 (6 to 12) vs. 10 (5 to 11) mm).

No leakage was observed at the stump in the pancreas in the ligation group (Fig. *6A*). However, in the stapler group, six of 10 pancreas showed leakage at the staple line or into the parenchyma (Fig. *6B*). The MPD pressure resistance was significantly higher in the ligation group than in the stapler group (median: 42.8 vs. 34.3 mmHg, P = 0.023) (Fig. 7).



#### FIGURE 6: Fluoroscopic imaging during the pressure resistance test.

(A) Fluoroscopic view with the pancreas ligation device. The device stops the flow of contrast medium, with no leakage detected from the pancreatic stump. (B) Fluoroscopic view after closure with the linear stapler. Leakage into the panenchyma is detected.



### FIGURE 7: Main pancreatic duct pressure resistance test results.

The pressure resistance is significantly higher in the ligation group than in the stapler group (P = 0.023).

# **Discussion**

In a previous pilot study [20], we introduced an innovative pancreas ligation device potentially able to

prevent POPF after distal pancreatectomy by atraumatic ligation. The device ligates the pancreatic stump while maintaining the arterial blood flow and limiting pancreatic necrosis. After modification of the ligation device, the present *ex vivo* follow-up study was conducted to evaluate the effectiveness of the improved ligation device by comparing the MPD pressure resistance with the improved ligation device against that with a regular linear stapler. The present study highlights an important finding. That is, compared to the regular linear stapler closure, this new device adequately ligated the pancreatic stump with a significantly higher MPD pressure resistance, thereby reducing the risk of POPF.

The normal MPD pressure among 20 healthy humans was estimated as a mean of 16.2 mmHg with a range of 3-31 mmHg [21,22]. A previous *ex vivo* study reported that the mean pressure associated with leakage after distal pancreatectomy was 13.7 mmHg when using 2.3-mm staples for stump closure [23]. These findings suggested that pancreatic leakage might occur within the normal MPD pressure in humans after distal pancreatectomy when using a routine linear stapler. Based on these reports, a pancreatic stump should be higher than 13.7 mmHg (the stapler's mean pressure) and ideally higher than 31 mmHg (the upper limit of pancreatic duct pressure in healthy humans) as described in the above previous studies. In the present study, all pancreases in the ligation group demonstrated no leakage at a median MPD pressure of 42.8 mmHg, which was significantly higher than that in the stapler group. Thus, our device can close the pancreatic stump with sufficient MPD pressure resistance to reduce the risk of POPF.

Surgical staplers are typically used to close the pancreatic stump in distal pancreatectomy. A multicenter randomized controlled trial (DISPACT trial) comparing stapler and hand-sewn closure of the pancreatic stump previously determined no significant difference in the occurrence of POPF with subsequent death after distal pancreatectomy until postoperative day seven [4]. Several new techniques and devices have been introduced to minimize POPF complications after distal pancreatectomy, such as pancreaticojejunostomy and reinforced staplers. However, a multicenter randomized controlled trial showed that pancreaticojejunostomy did not significantly reduce the occurrence of POPF after distal pancreatectomy compared to conventional surgical stapler closure [24]. Although some studies have reported the superiority of reinforced staplers over regular stapler closure [25,26], the occurrence of clinically relevant POPF has remained high at 5-16% [25-27]. In multicenter randomized clinical trials, reinforced staplers did not show any advantage over regular staplers for clinically relevant POPF prevention [17,18]. These conventional methods cannot close the pancreatic stump without penetrating the pancreatic parenchyma with staples or needles. We hypothesized that such injuries would lead to pancreatic duct and small artery injury at the pancreatic stump.

Our previous report [20] showed that ischemia of the pancreatic stump is associated with pancreatic necrosis. Nagakawa et al. [23] reported that hand-sewn closure caused pancreatic necrosis and regenerated the pancreatic ducts in mongrel dogs, potentially leading to POPF. These reports suggest that maintaining the arterial blood flow to the pancreatic stump may reduce the risk of necrosis and incidence of POPF. Yamashita et al. [28] developed a clip-type device to reduce injury to the pancreatic parenchyma during stump closure. This clip has a space between the upper and lower jaws to avoid causing direct injury to the pancreatic parenchyma. However, the clip may not control the ligation force, thereby less controlling the arterial blood flow to the pancreatic stump. An ideal device for pancreatic stump closure can minimize the risk of pancreatic duct and artery injury during ligation while maintaining the arterial blood flow to the pancreatic addresses these previous issues and may also reduce the incidence of POPF.

The present study has some limitations. First, a porcine pancreas was used as a substitute for the human pancreas to assess leak pressure. Our findings may therefore not be directly generalizable to human clinical practice. However, the porcine pancreas has soft parenchyma, similar to the human pancreas, and has been used in studies of distal pancreatectomy [29,30]. Second, the study did not assess any ischemic changes at the pancreatic stump or ligation site because of the *ex vivo* experimental study design. The ligation pressure in this study might affect arterial blood flow reduction to the pancreatic stump. Although the band-type ligation device [20] could preserve arterial blood flow to the pancreatic stump and prevent POPF, future *in vivo* studies using a front-to-back clamping-type device are warranted to confirm the preservation of arterial blood flow.

# Conclusions

The present study demonstrates the effectiveness of an innovative pancreas ligation device in reducing the incidence of POPF after distal pancreatectomy. Compared to the regular linear stapler closure, the device ligated the pancreatic stump with adequate pressure to avoid leakage. Our ligation device is expected to be a useful alternative to a linear stapler for pancreatic stump closure.

# **Additional Information**

#### Disclosures

Human subjects: All authors have confirmed that this study did not involve human participants or tissue. Animal subjects: The Animal Experiment Committee Issued protocol number 18035-02. Conflicts of **interest:** In compliance with the ICMJE uniform disclosure form, all authors declare the following: **Payment/services info:** Conflicts of interest/competing interests: Yuji Kaneda. Funding: This work was financially supported by Grants-in-Aid for Scientific Research from the Japan Society for the Promotion of Science (grant number JP20K17701) and the Japan Agency for Medical Research and Development (AMED) (grant number JP20Im0203003) and grants-in-aid from the 106th Annual Congress of JSS Memorial Surgical Research Fund, Tokyo, Japan (grant number 2019-1-07) and Teijin Medical Technologies Co., Ltd., Osaka, Japan. **Financial relationships:** All authors have declared that they have no financial relationships at present or within the previous three years with any organizations that might have an interest in the submitted work. **Intellectual property info:** United States Patent; patent no.: US 11337704; date of patent: May 24, 2022; Tool for Treating Excised End of Body Organ; applicant: Jichi Medical University, Tokyo (JP); inventor: Yuji Kaneda, Shimotsuke (JP). **Other relationships:** All authors have declared that there are no other relationships or activities that could appear to have influenced the submitted work.

#### Acknowledgements

We thank all of the staff at the Center for Development of Advanced Medical Technology, Jichi Medical University, for their technical support in our animal experiments.

## References

- 1. Sledzianowski JF, Duffas JP, Muscari F, Suc B, Fourtanier F: Risk factors for mortality and intra-abdominal morbidity after distal pancreatectomy. Surgery. 2005, 137:180-5. 10.1016/j.surg.2004.06.063
- Kleeff J, Diener MK, Z'graggen K, et al.: Distal pancreatectomy: risk factors for surgical failure in 302 consecutive cases. Ann Surg. 2007, 245:573-82. 10.1097/01.sla.0000251438.43135.fb
- Harris LJ, Abdollahi H, Newhook T, et al.: Optimal technical management of stump closure following distal pancreatectomy: a retrospective review of 215 cases. J Gastrointest Surg. 2010, 14:998-1005. 10.1007/s11605-010-1185-z
- Diener MK, Seiler CM, Rossion I, et al.: Efficacy of stapler versus hand-sewn closure after distal pancreatectomy (DISPACT): a randomised, controlled multicentre trial. Lancet. 2011, 377:1514-22. 10.1016/s0140-6736(11)60237-7
- Goh BK, Tan YM, Chung YF, et al.: Critical appraisal of 232 consecutive distal pancreatectomies with emphasis on risk factors, outcome, and management of the postoperative pancreatic fistula: a 21-year experience at a single institution. Arch Surg. 2008, 143:956-65. 10.1001/archsurg.143.10.956
- Yanagimoto H, Satoi S, Toyokawa H, et al.: Pancreaticogastrostomy following distal pancreatectomy prevents pancreatic fistula-related complications. J Hepatobiliary Pancreat Sci. 2014, 21:473-8. 10.1002/jhbp.59
- Ma LW, Dominguez-Rosado I, Gennarelli RL, et al.: The cost of postoperative pancreatic fistula versus the cost of pasireotide: results from a prospective randomized trial. Ann Surg. 2017, 265:11-6. 10.1097/SLA.000000000001892
- Merkow RP, Bilimoria KY, Tomlinson JS, et al.: Postoperative complications reduce adjuvant chemotherapy use in resectable pancreatic cancer. Ann Surg. 2014, 260:372-7. 10.1097/SLA.00000000000378
- Sandini M, Ruscic KJ, Ferrone CR, et al.: Major complications independently increase long-term mortality after pancreatoduodenectomy for cancer. J Gastrointest Surg. 2019, 23:1984-90. 10.1007/s11605-018-3939-y
- Nagai S, Fujii T, Kodera Y, et al.: Recurrence pattern and prognosis of pancreatic cancer after pancreatic fistula. Ann Surg Oncol. 2011, 18:2329-37. 10.1245/s10434-011-1604-8
- 11. Hank T, Sandini M, Ferrone CR, et al.: Association between pancreatic fistula and long-term survival in the era of neoadjuvant chemotherapy. JAMA Surg. 2019, 154:943-51. 10.1001/jamasurg.2019.2272
- Bassi C, Marchegiani G, Dervenis C, et al.: The 2016 update of the International Study Group (ISGPS) definition and grading of postoperative pancreatic fistula: 11 years after. Surgery. 2017, 161:584-91. 10.1016/j.surg.2016.11.014
- Eshmuminov D, Schneider MA, Tschuor C, et al.: Systematic review and meta-analysis of postoperative pancreatic fistula rates using the updated 2016 International Study Group Pancreatic Fistula definition in patients undergoing pancreatic resection with soft and hard pancreatic texture. HPB (Oxford). 2018, 20:992-1003. 10.1016/j.hpb.2018.04.003
- Ecker BL, McMillan MT, Allegrini V, et al.: Risk factors and mitigation strategies for pancreatic fistula after distal pancreatectomy: analysis of 2026 resections from the international, multi-institutional distal pancreatectomy study group. Ann Surg. 2019, 269:143-9. 10.1097/SLA.00000000002491
- Chong E, Ratnayake B, Lee S, et al.: Systematic review and meta-analysis of risk factors of postoperative pancreatic fistula after distal pancreatectomy in the era of 2016 International Study Group pancreatic fistula definition. HPB (Oxford). 2021, 23:1139-51. 10.1016/j.hpb.2021.02.015
- van der Heijde N, Lof S, Busch OR, et al.: Incidence and impact of postoperative pancreatic fistula after minimally invasive and open distal pancreatectomy. Surgery. 2022, 171:1658-64. 10.1016/j.surg.2021.11.009
- Wennerblom J, Ateeb Z, Jönsson C, et al.: Reinforced versus standard stapler transection on postoperative pancreatic fistula in distal pancreatectomy: multicentre randomized clinical trial. Br J Surg. 2021, 108:265-70. 10.1093/bjs/znaa113
- Merdrignac A, Garnier J, Dokmak S, et al.: Effect of the use of reinforced stapling on the occurrence of pancreatic fistula after distal pancreatectomy: results of the REPLAY (REinforcement of the Pancreas in distaL pAncreatectomY) multicenter randomized clinical trial. Ann Surg. 2022, 276:769-75. 10.1097/SLA.000000000005618
- Miao Y, Lu Z, Yeo CJ, et al.: Management of the pancreatic transection plane after left (distal) pancreatectomy: Expert consensus guidelines by the International Study Group of Pancreatic Surgery (ISGPS). Surgery. 2020, 168:72-84. 10.1016/j.surg.2020.02.018
- 20. Kaneda Y, Kimura Y, Saito A, et al.: Innovative pancreas ligation band for distal pancreatectomy: a pilot in

vivo porcine study. Cureus. 2021, 13:e18238. 10.7759/cureus.18238

- Okazaki K, Yamamoto Y, Ito K: Endoscopic measurement of papillary sphincter zone and pancreatic main ductal pressure in patients with chronic pancreatitis. Gastroenterology. 1986, 91:409-18. 10.1016/0016-5085(86)90576-7
- Okazaki K, Yamamoto Y, Kagiyama S, et al.: Pressure of papillary sphincter zone and pancreatic main duct in patients with chronic pancreatitis in the early stage. Scand J Gastroenterol. 1988, 23:501-7. 10.3109/00365528809093901
- Nagakawa Y, Hijikata Y, Osakabe H, et al.: Why does postoperative pancreatic fistula occur after hand-sewn parenchymal closure and staple closure in distal pancreatectomy?. Surg Laparosc Endosc Percutan Tech. 2019, 29:e15-9. 10.1097/SLE.00000000000604
- 24. Kawai M, Hirono S, Okada K, et al.: Randomized controlled trial of pancreaticojejunostomy versus stapler closure of the pancreatic stump during distal pancreatectomy to reduce pancreatic fistula. Ann Surg. 2016, 264:180-7. 10.1097/SLA.00000000001395
- 25. Yamashita YI, Tsujita E, Chikamoto A, et al.: Linear stapling device with pre-attached bioabsorbable polyglycolic acid felt reduces postoperative pancreatic fistula after distal pancreatectomy. Anticancer Res. 2017, 37:1865-8. 10.21873/anticanres.11523
- Kondo N, Uemura K, Nakagawa N, et al.: A multicenter, randomized, controlled trial comparing reinforced staplers with bare staplers during distal pancreatectomy (HiSCO-07 Trial). Ann Surg Oncol. 2019, 26:1519-27. 10.1245/s10434-019-07222-0
- Kawai M, Hirono S, Okada KI, et al.: Reinforced staplers for distal pancreatectomy. Langenbecks Arch Surg. 2017, 402:1197-204. 10.1007/s00423-017-1634-1
- Yamashita YI, Yamao T, Nakao Y, et al.: Efficacy of a newly developed bioabsorbable pancreatic clip for distal pancreatectomy in swine. Surg Today. 2022, 52:1109-14. 10.1007/s00595-021-02435-x
- Hornez E, Garnier E, Sastre B, Garcia S, Mayet A, Berdah SV: Bioabsorbable staple-line reinforcement for pancreatectomy in a porcine model: a preliminary study. Eur Surg Res. 2012, 48:48-53. 10.1159/000333397
- Wang D, Ji ZL, Jiang XH, Wang JM, Tan YY, Wang Y, Wen YZ: Laparoendoscopic single-site distal pancreatectomy in pigs. World J Gastroenterol. 2014, 20:6878-83. 10.3748/wjg.v20.i22.6878