Valve-in-Valve Transcatheter Aortic Valve Implantation Using Fluoroscopic Guidance for Late Failure of the Bioprosthetic Bentall Conduit BioValsalva™ Vascutek

Pedro Teixeira¹, Wilson Ferreira¹, Gustavo Pires-Morais¹, Alberto Rodrigues², Pedro Braga¹

1. Cardiology, Centro Hospitalar de Vila Nova de Gaia / Espinho, Vila Nova de Gaia, PRT 2. Cardiology, Centro Hospitalar de Vila Nova de Gaia / Espinho, Vila Nova De Gaia, PRT

Corresponding author: Pedro Teixeira, pedromgoncalvesteixeira@gmail.com

Abstract

We describe one of the first-in-human cases of valve-in-valve transcatheter aortic valve implantation (TAVI) using fluoroscopic-only guidance for the treatment of late failure of a bio-prosthetic Bentall conduit, the BioValsalva™ Vascutek (Vascutek Terumo, Renfrewshire, Scotland), using a self-expandable heart valve prosthesis (CoreValve™ Evolut™ R, Medtronic, Dublin Ireland).

Categories: Cardiac/Thoracic/Vascular Surgery, Cardiology

Keywords: transcatheter aortic valve implantation, valve-in-valve, bio-prosthetic conduit failure

Introduction

Over the past decade, transcatheter aortic valve implantation (TAVI) has emerged as a valid and safe treatment option for severe symptomatic aortic stenosis, from prohibitive surgical risk patients in the early days to those at low-to-intermediate risk in current practice [1-6].

Advances in device technology, as well as growing operators' skills, were remarkable in this short time window. This set the stage for a fantastic expansion of the procedure indications towards increasingly challenging clinical scenarios. Valve-in-valve TAVI has become a safe and reproducible procedure for patients with severe bioprosthetic valve dysfunction deemed to be at high risk for re-operation.

Pure (native or prosthetic) valve regurgitation is still a grey zone, and evidence in this challenging subset is much needed. These patients typically exhibit less annular and/or leaflet calcifications, implying technical difficulties in adequate transcatheter aortic valve anchoring. Furthermore, non-calcified native aortic valves and non-radiopaque bioprosthesis both challenge correct positioning under fluoroscopic guidance.

We describe one of the first in-human cases of valve-in-valve TAVI using fluoroscopic-only guidance for the treatment of late failure of a bio-prosthetic Bentall conduit, the BioValsalva™ Vascutek (Vascutek Terumo, Renfrewshire, Scotland), using a self-expandable heart valve prosthesis (CoreValve™ Evolut™ R, Medtronic, Dublin Ireland).

Case Presentation

The patient was a 65-year-old woman with no known cardiovascular risk factors. Relevant past medical history included epilepsy and primary biliary cirrhosis. She was submitted to a Bentall surgical procedure, with implantation of a 25mm porcine bioprosthetic valved conduit (BioValsalva™ Vascutek) in 2008, following documentation of bicuspid aortic valve with severe regurgitation and a dilated ascending aorta.

She was chronically medicated with venlafaxine, levetiracetam, pantoprazole, aspirin, furosemide, and bisoprolol.

The patient presented with new-onset heart failure symptoms twelve-years after the Bentall procedure, New York Heart Association (NYHA) III functional class, and severe transprosthetic aortic regurgitation was diagnosed. She was refused for re-do cardiac surgery considering high surgical risk - the calculated EuroScore II was 18.52%, and the Society of Thoracic Surgeons (STS) score for mortality was 8.60%.

Procedural planning

Transoesophageal echocardiogram showed thickness and retraction of the valve cusp positioned in the right coronary position and a prolapse of the cusp in the non-coronary position resulting in severe, eccentric, transprosthetic aortic regurgitation. There was also a mild enlargement of the left cardiac chambers

Review began 12/01/2020

Review ended 12/05/2020

Published 12/10/2020

© Copyright 2020

Teixeira et al. This is an open access article distributed under the terms of the Creative Commons Attribution License CC-BY 4.0., which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

How to cite this article

Teixeira P, Ferreira W, Pires-Morais G, et al. (December 10, 2020) Valve-in-Valve Transcatheter Aortic Valve Implantation Using Fluoroscopic Guidance for Late Failure of the Bioprosthetic Bentall Conduit BioValsalva™ Vascutek. Cureus 12(12): e12006. DOI 10.7759/cureus.12006

(indexed left ventricular end-diastolic volume of 68ml/m²), and left ventricular systolic function was mildly depressed (left ventricle ejection fraction of 49%). Coronary artery disease was excluded by invasive angiography. A pre-procedure CT scan is shown in figure 1.

FIGURE 1: Pre-procedure CT scan

LCA - left coronary artery; LVOT - left ventricle outflow tract; RCA - right coronary artery; STJ - sino-tubular junction

Detailed aortic annulus and aortic root measurements are shown in Table 1. Luminal narrowing, near the neo-Valsalva sinuses, introduced additional planning difficulties (Figure 2), and the measurement rendered a 23mm CoreValve as the most suitable prosthetic valve. Although not extensively, small calcium foci were present at the leaflets and the sinotubular level of the conduit. No significant calcification was present either at the annulus or left ventricular outflow tract level.

Parameter	Value
Aortic annulus	
Perimeter	56.4 mm
Area	254 mm ²
Maximum dimension	18.9 mm
Minimum dimension	16.9 mm
LCA distance	10.8 mm
RCA distance	11.3 mm
Neo-Valsalva sinus	
Diameter	27 x 26 x 26 mm

TABLE 1: Aortic annulus and aortic root measurements

LCA - left coronary artery; RCA - right coronary artery

FIGURE 2: Pre-procedure three dimensional (3D) CT scan reconstruction, showing a luminal narrowing near the neo-Valsalva sinuses, which created measurement and planning difficulties

TAVI procedure

The procedure was performed using right transfemoral access, under fluoroscopic guidance. Baseline aortography (Figure 3, left upper panel) shows severe eccentric aortic regurgitation, while it also provided coronary's patency and height evaluation. Balloon pre-dilatation was performed with a 20mm TrueDilatation™ balloon (Figure 3, right upper panel). Pre-dilatation was considered to provide additional reassurance in estimating the internal diameter of this poorly documented bioprosthesis model. It also permitted to give support to the chosen prosthesis size, which was not oversized a priori. Furthermore, adequate coronary opacification confirmed the safety of the planned implantation site. Calcium foci present at the leaflets level proved useful at this stage, helping accurate determination of the implantation depth.

FIGURE 3: Fluoroscopic guidance during TAVI procedure; LAO 12 CRAN 5 projection was used

TAVI - transcatheter aortic valve implantation; CRAN - cranial; LAO - left anterior oblique

A self-expandable valve prosthesis was then implanted at this stage, CoreValve Evolut R 23mm, under rapid ventricular pacing (Figure 3, left lower panel), followed by post-dilatation using the 20mm (TrueDilatation) balloon (Figure 3, right lower panel). The final angiography showed mild aortic regurgitation. There was no significant transaortic gradient by the end of the procedure. Vascular access site closure was uneventful following two suture-based pre-closure devices (ProGlide™). No electrocardiographic changes were observed in the peri-procedural period.

Pre-discharge transthoracic echocardiogram showed biologic aortic valve prosthesis with maximum and mean gradients of 23 and 12mmHg, respectively. Mild to moderate paravalvular leak between the percutaneous aortic valve and the bioprosthetic valved conduit was visible (Figure 4).

FIGURE 4: Pre-discharge transthoracic echocardiogram

A parasternal short-axis view at the aortic valve level is shown in the left panel. An apical long-axis view is shown in the right panel.

After four days, the patient was discharged home, and a significant symptomatic improvement was documented in early follow-up (NYHA class II, two weeks following the procedure).

Discussion

A recently published meta-analysis comprising 594 patients from six observational studies, of whom 255 underwent valve-in-valve (ViV) TAVI and 339 underwent redo surgical aortic valve replacement (SAVR), reiterated the safety and feasibility of ViV-TAVI for failed aortic bio-prosthetic valves in patients deemed to be at high risk for surgery [7]. In this analysis, no significant difference between groups was observed for procedural, 30 days, and one-year mortality rates. Valve-in-valve TAVI was associated with a lower risk of permanent pacemaker implantation with a trend toward increased risk of paravalvular leak.

Data from the Global Valve-in-Valve Registry showed that bioprosthetic valve failure might present as predominant regurgitation in up to 34% of patients, as a consequence of wear and tear, calcification, or infection [8]. ViV-TAVI was performed with the Edwards SAPIEN (Edwards Lifesciences, Irvine, USA) or the CoreValve (Medtronic, Minneapolis, USA) devices. Although stentless aortic valves represented 23% of patients in this large registry, from which 30% were homografts, the literature is scarce concerning ViV-TAVI for failed bio-prosthetic aortic conduits, predominantly due to regurgitation.

Another group reported successful treatment of a failed BioConduit with severe regurgitation, using transoesophageal echocardiographic guidance [9]. ViV TAVI after Bentall operation with a homograft was also previously reported for the treatment of severe bioprosthetic stenosis [10].

Conclusions

This is, to the best of our knowledge, one of the first literature reports of a valve-in-valve TAVI for the treatment of late failure of a BioValsalva Vascutek conduit with severe aortic regurgitation, using the CoreValve device. The patient had an uneventful recovery in the early postoperative period and was discharged after four days. With this challenging case, the authors show the feasibility and safety of ViV-TAVI in this innovative scenario, when performed in high-volume centers and by experienced operators.

Additional Information

Disclosures

Human subjects: Consent was obtained by all participants in this study. **Conflicts of interest:** In compliance with the ICMJE uniform disclosure form, all authors declare the following: **Payment/services**

info: All authors have declared that no financial support was received from any organization for the submitted work. **Financial relationships:** All authors have declared that they have no financial relationships at present or within the previous three years with any organizations that might have an interest in the submitted work. **Other relationships:** All authors have declared that there are no other relationships or activities that could appear to have influenced the submitted work.

References

1. Leon MB, Smith CR, Mack M, et al.: Transcatheter aortic-valve implantation for aortic stenosis in patients who cannot undergo surgery. *N Engl J Med*. 2010, 363:1597-1607. [10.1056/NEJMoa1008232](https://doi.org/10.1056/NEJMoa1008232)
2. Adams DH, Popma JJ, Reardon MJ, et al.: Transcatheter aortic-valve replacement with a self-expanding prosthesis. *N Engl J Med*. 2014, 370:1790-1798. [10.1056/NEJMoa1400590](https://doi.org/10.1056/NEJMoa1400590)
3. Leon MB, Smith CR, Mack M, et al.: Transcatheter or surgical aortic-valve replacement in intermediate-risk patients. *N Engl J Med*. 2016, 374:1609-1620. [10.1056/NEJMoa1514616](https://doi.org/10.1056/NEJMoa1514616)
4. Reardon MJ, Van Mieghem NM, Popma JJ, et al.: Surgical or transcatheter aortic-valve replacement in intermediate-risk patients. *N Engl J Med*. 2017, 376:1321-1331. [10.1056/NEJMoa1700456](https://doi.org/10.1056/NEJMoa1700456)
5. Mack MJ, Leon MB, Thourani VH, et al.: Transcatheter aortic-valve replacement with a balloon-expandable valve in low-risk patients. *N Engl J Med*. 2019, 380:1695-1705. [10.1056/NEJMoa1814052](https://doi.org/10.1056/NEJMoa1814052)
6. Popma JJ, Deeb GM, Yakubov SJ, et al.: Transcatheter aortic-valve replacement with a self-expanding valve in low-risk patients. *N Engl J Med*. 2019, 380:1706-1715. [10.1056/NEJMoa1816885](https://doi.org/10.1056/NEJMoa1816885)
7. Nalluri N, Atti V, Munir AB, et al.: Valve in valve transcatheter aortic valve implantation (ViV-TAVI) versus redo—Surgical aortic valve replacement (redo-SAVR): A systematic review and meta-analysis. *J Interv Cardiol*. 2018, 31:661-671. [10.1111/joic.12520](https://doi.org/10.1111/joic.12520)
8. Dvir D, Webb J, Brecker S, et al.: Transcatheter aortic valve replacement for degenerative bioprosthetic surgical valves: results from the Global Valve-in-Valve Registry. *Circulation*. 2012, 126:2335-2344. [10.1161/CIRCULATIONAHA.112.104505](https://doi.org/10.1161/CIRCULATIONAHA.112.104505)
9. Budassi S, Verheye V, Agostoni P, et al.: Valve-in-valve transcatheter aortic valve replacement in a patient with late failure of a bioprosthetic Bentall procedure. *Struct Heart*. 2020, 4:236-239. [10.1080/24748706.2020.1742405](https://doi.org/10.1080/24748706.2020.1742405)
10. Joo HJ, Hong SJ, Yu CW: Transcatheter aortic valve-in-valve implantation for severe bioprosthetic stenosis after Bentall operation using a homograft in a patient with Behçet's disease. *J Heart Valve Dis*. 2015, 24:217-219.