Review began 01/20/2022
Review ended 01/27/2022
Published 02/03/2022

© Copyright 2022

Bhullar et al. This is an open access article distributed under the terms of the Creative Commons Attribution License CC-BY 4.0., which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

Comparative Analysis of Axial Length Measurements by Optical Biometers Based on Partial Coherence Interferometry Versus Optical Low-Coherence Interferometry: An Office Audit

Harleen Bhullar¹, Leland Dhurjon², Christopher Francis², Vishaal Bhambhwani^{2,3}

1. Medicine, American University of the Caribbean School of Medicine, Cupecoy, SXM 2. Department of Surgery, Thunder Bay Regional Health Sciences Centre, Thunder Bay, CAN 3. Department of Surgery, Northern Ontario School of Medicine, Thunder Bay, CAN

Corresponding author: Vishaal Bhambhwani, vishaalb@gmail.com

Abstract

Purpose

In this study, we aimed to compare axial length (AL) measurements of the IOLMaster 500 (Carl Zeiss Meditec AG, Jena, Germany), based on partial coherence interferometry (PCI) versus the Aladdin (Topcon Healthcare, Oakland, NJ), based on optical low-coherence interferometry (OLCI), in a clinical setting.

Methods

A retrospective analysis of the records of patients presenting for cataract surgery at an ophthalmology practice between October 2019 and March 2020 was performed. All patients had biometry measurements on the IOLMaster 500 and the Aladdin. Data collected included patient demographics, cataract morphology and type, and AL measurements. The IOLMaster 500 and Aladdin measurements were compared using the unpaired t-test and Chi-squared test.

Results

In total, 393 eyes (197 patients) were included (91 males, 107 females) in the study. The IOLMaster 500 was unable to successfully obtain AL measurements in seven eyes (1.8%) and the Aladdin in 26 eyes (6.6%). The difference was statistically significant ($p=0.0007$). Advanced and central posterior subcapsular cataracts were common in eyes that had unsuccessful measurements. In the eyes successfully measured, the mean AL for the IOLMaster was 24.04 ± 1.32 mm, while it was 24.04 ± 1.34 mm for the Aladdin. However, this difference was not statistically significant ($p=0.9165$).

Conclusion

The IOLMaster 500 performed better in terms of the number of eyes for which AL measurements were successfully obtained compared to the Aladdin. This may be partly related to high volumes of advanced cataracts treated at our practice. However, this being a retrospective study, a cause-and-effect relationship could not be established.

Categories: Ophthalmology

Keywords: cataract, optical low coherence interferometry, partial coherence interferometry, optical biometers, axial length, biometry

Introduction

Cataract surgery with the implantation of an intraocular lens (IOL) is one of the most commonly performed ophthalmic surgical procedures [1,2]. To optimize the accuracy of predicting postoperative refraction, formulas have been developed to calculate IOL power on the basis of preoperative measurement of intraocular distances, especially axial length (AL) and keratometry [3,4]. Previously, applanation and immersion ultrasound (contact methods) were used for AL measurements [5]. Later on, optical (non-contact) biometry techniques were developed [6].

Until 2009, the IOLMaster (Carl Zeiss Meditec AG, Jena, Germany), based on partial coherence interferometry (PCI) with a 780 nm diode laser, had been the only device to measure AL by the optical method [7]. Since then, newer instruments have been introduced. One of the newer biometers is the Aladdin (Topcon Healthcare, Oakland, NJ), based on optical low-coherence interferometry (OLCI) using an 830 nm super-luminescent diode [7,8].

A few studies comparing the IOLMaster with the Aladdin are available in the literature [7-10]; however, most

How to cite this article

Bhullar H, Dhurjon L, Francis C, et al. (February 03, 2022) Comparative Analysis of Axial Length Measurements by Optical Biometers Based on Partial Coherence Interferometry Versus Optical Low-Coherence Interferometry: An Office Audit. Cureus 14(2): e21883. DOI 10.7759/cureus.21883

were performed in a controlled experimental environment rather than in real-life clinical settings. The purpose of this study is to compare and contrast the performance of the IOLMaster 500 with the Aladdin using real-life clinical data from an office audit.

The results from this study were presented as a free paper at the European Society of Cataract and Refractive Surgeons (ESCRS) 2021 Virtual Winter Meeting, held from 19th to 21st February 2021.

Materials And Methods

A retrospective analysis of records of patients who presented for cataract surgery at an ophthalmology office in Northern Ontario, Canada between October 2019 and March 2020 was performed. The office is shared by three ophthalmologists and houses both the IOLMaster 500 (v. 7.7) and the Aladdin (v. 1.1.3) biometers. All patients who present for cataract surgery at the office get their measurements on both these biometers. Biometry measurements are performed and recorded by trained ophthalmology technicians during the course of the patient visits. Patients also undergo detailed clinical evaluation by ophthalmologists during their visit, which is recorded in their charts.

Based on the analysis of the patient charts, the data collected for the purpose of the study included patient demographics (age, sex), clinical exam findings (affected eye/s, cataract morphology, and type), and AL measurements recorded by the two biometers (the IOLMaster 500 and the Aladdin). The number of eyes for which AL measurements were successfully obtained by the IOLMaster 500 and the Aladdin was recorded. These were analyzed and compared using the Chi-squared test. For eyes with successful measurements, mean ALs in the two biometer groups were calculated and the unpaired t-test was used for statistical analysis. For both comparisons, a p-value of <0.05 was considered statistically significant. For eyes with unsuccessful measurements, cataract morphology and type were analyzed with a view to determining the possible causes of the biometer's failure to obtain a measurement. The study received a waiver from the institutional Ethics Review Board.

Results

A total of 197 patients (393 eyes) were included in the study. There were 91 males and 106 females. The mean age of the patients was 71.76 ±8.45 years (range: 40-94 years). Table 1 and Table 2 show the AL measurements recorded by the IOLMaster 500 and the Aladdin.

	Axial length by the Aladdin	Axial length by the IOLMaster
Total number of eyes	393	393
Measured/recorded	367	386
Not measured/not recorded	26	7
P-value (Chi-squared test)	0.000727	

TABLE 1: Axial length measurements (successful and unsuccessful) by the two biometers

	Axial length (mm) (mean ±standard deviation)
IOLMaster	24.04 ±1.32
Aladdin	24.04 ±1.34
P-value (unpaired t-test)	0.9165

TABLE 2: Axial length measurements (mean and standard deviation) by the two biometers

As shown in Table 1, the IOLMaster 500 was unable to obtain measurements of AL in seven eyes and the Aladdin in 26 eyes. The difference was statistically significant (1.8% versus 6.6%, p=0.0007). Table 2 shows the mean values (with standard deviation) for AL between the two biometers for eyes that were successfully measured/recorded. In eyes successfully measured, the mean AL for the IOLMaster was 24.04±1.32 mm, while it was 24.04 ±1.34 mm for the Aladdin. The difference was not statistically significant (p=0.9165).

For eyes in which AL was not successfully measured by either of the two biometers, patient charts were

reviewed to find morphology and type of cataract (based on the World Health Organization Simplified Cataract Grading System, 2002) (Table 3). As seen in Table 3, among the eyes with unmeasured AL, a large percentage had advanced or central cataracts.

Cataract morphology and type	Number of eyes (%)
Nuclear sclerosis grade	
3	2 (7%)
4	7 (25%)
Posterior subcapsular cataract grade	
2	4 (14%)
3	4 (14%)
Anterior subcapsular cataract	
Present	2 (7%)
Mature or total cataract	
Present	4 (25%)

TABLE 3: Cataract morphology and type among eyes with unsuccessful axial length measurements

Discussion

Our study analyzed real-life clinical data from an office audit to compare two different instruments and technologies for optical biometry (the IOLMaster 500 based on PCI versus the Aladdin based on OLCI); to the best of our knowledge, this is one of the largest studies of this nature.

Accurate and repeatable biometry measurements are of utmost importance for ideal post-cataract surgery refractive outcomes [3,4]. Our study shows that the IOLMaster 500 performed better than the Aladdin in terms of the number of eyes for which measurements of AL were successfully obtained. Thus, IOL power calculation was possible for more eyes with the IOLMaster 500 compared to the Aladdin, in our practice.

Our study also attempted to explore the relationship between unsuccessful measurement of AL by the two biometers and the morphology and type of cataract. Being a retrospective study, a cause-and-effect relationship could not be established. However, eyes with unmeasured AL had a high percentage of advanced, mature/total, and posterior subcapsular cataracts, which were likely responsible for the unsuccessful measurements [11,12]. Our practice serves as a referral center that caters to a large and scattered population in Northern Ontario. This is a region with a vast geographic extent and significant physician shortages, where a lot of patients have to travel long distances for access to cataract surgery [13,14]. Our cataract practice hence deals with a lot of patients with advanced cataracts, which may explain the difference in successful measurements between the two biometers. In contrast to some of the previously published literature [9], we hypothesize that PCI may be better than OLCI in obtaining successful measurements in eyes with advanced and central cataractous changes. As reported previously in the literature [7-10], our study reiterates that AL measurements by the two biometers show no statistically significant difference, as long as the machine successfully measures these parameters.

There are a few limitations to our study. Being a retrospective study, a direct cause-and-effect relationship for unmeasured AL values could not be determined. It is possible that the same technician may not have performed both tests on one patient. Also, since our office had acquired the IOLMaster 500 before the Aladdin, it may be possible that our technicians are more adept at using the IOLMaster. However, the relative ease and rapidity of the Aladdin measurements [8] may make technician training and skills less significant [15].

We are aware from the published literature that both the IOLMaster 500 and the Aladdin have their distinct advantages and disadvantages [7-10]. Hence, a physician’s scope of practice (especially the morphology and type of cataracts the practice commonly encounters) may be the deciding factor regarding the preferred biometer choice.

Conclusions

For patients undergoing cataract surgery, accurate measurement of AL is of utmost importance for accurate IOL power calculation, so as to achieve optimal postoperative refraction. Our study shows that the IOLMaster 500 and the Aladdin biometers provide comparable AL measurements (as long as the measurements are successfully obtained). This is in agreement with the previously published literature. However, in our study, the IOLMaster 500 was significantly better in terms of the number of eyes for which successful measurements of AL were obtained. Thus, IOL power calculation was possible in more eyes with the IOLMaster 500 compared to the Aladdin, in our study.

One of the factors responsible for this difference may be the high numbers of advanced and central cataracts treated at our practice, possibly due to its geographic location with its associated local population and health demographic. Thus, we conclude that the type of patients and cataracts that a practice caters to may be an important factor in determining the biometer choice. For ophthalmology practices that treat a large number of advanced cataracts and posterior subcapsular cataracts, the IOLMaster 500 (based on PCI) may be better suited than the Aladdin (based on OLCI).

Additional Information

Disclosures

Human subjects: Consent was obtained or waived by all participants in this study. Thunder Bay Regional Health Sciences Centre Research Ethics Office issued approval N/A. This project falls under the TCPS2 Chapter 2, Article 2.5 as a quality improvement study and does not require research ethics approval. **Animal subjects:** All authors have confirmed that this study did not involve animal subjects or tissue. **Conflicts of interest:** In compliance with the ICMJE uniform disclosure form, all authors declare the following: **Payment/services info:** All authors have declared that no financial support was received from any organization for the submitted work. **Financial relationships:** All authors have declared that they have no financial relationships at present or within the previous three years with any organizations that might have an interest in the submitted work. **Other relationships:** All authors have declared that there are no other relationships or activities that could appear to have influenced the submitted work.

Acknowledgements

Yrena Malo, Kimberley Francis, and Nancy Oracz at our practice location provided assistance with data collection for the study.

References

1. Liu YC, Wilkins M, Kim T, Malyugin B, Mehta JS: Cataracts. *Lancet*. 2017, 390:600-12. [10.1016/S0140-6736\(17\)30544-5](https://doi.org/10.1016/S0140-6736(17)30544-5)
2. Lee CM, Afshari NA: The global state of cataract blindness. *Curr Opin Ophthalmol*. 2017, 28:98-103. [10.1097/ICU.0000000000000340](https://doi.org/10.1097/ICU.0000000000000340)
3. Norrby S: Sources of error in intraocular lens power calculation. *J Cataract Refract Surg*. 2008, 34:368-76. [10.1016/j.jcrs.2007.10.031](https://doi.org/10.1016/j.jcrs.2007.10.031)
4. Olsen T: Calculation of intraocular lens power: a review. *Acta Ophthalmol Scand*. 2007, 85:472-85. [10.1111/j.1600-0420.2007.00879.x](https://doi.org/10.1111/j.1600-0420.2007.00879.x)
5. Schelenz J, Kammann J: Comparison of contact and immersion techniques for axial length measurement and implant power calculation. *J Cataract Refract Surg*. 1989, 15:425-8. [10.1016/s0886-3350\(89\)80062-8](https://doi.org/10.1016/s0886-3350(89)80062-8)
6. Santodomingo-Rubido J, Mallen EA, Gilmartin B, Wolffsohn JS: A new non-contact optical device for ocular biometry. *Br J Ophthalmol*. 2002, 86:458-62. [10.1136/bjo.86.4.458](https://doi.org/10.1136/bjo.86.4.458)
7. Hoffer KJ, Shammas HJ, Savini G, Huang J: Multicenter study of optical low-coherence interferometry and partial-coherence interferometry optical biometers with patients from the United States and China. *J Cataract Refract Surg*. 2016, 42:62-7. [10.1016/j.jcrs.2015.07.041](https://doi.org/10.1016/j.jcrs.2015.07.041)
8. Mandal P, Berrow EJ, Naroo SA, Wolffsohn JS, Uthoff D, Holland D, Shah S: Validity and repeatability of the Aladdin ocular biometer. *Br J Ophthalmol*. 2014, 98:256-8. [10.1136/bjophthalmol-2013-304002](https://doi.org/10.1136/bjophthalmol-2013-304002)
9. Ortiz A, Galvis V, Tello A, Viana V, Corrales MI, Ochoa M, Rodriguez CJ: Comparison of three optical biometers: IOLMaster 500, Lenstar LS 900 and Aladdin. *Int Ophthalmol*. 2019, 39:1809-18. [10.1007/s10792-018-1006-z](https://doi.org/10.1007/s10792-018-1006-z)
10. Sabatino F, Findl O, Maurino V: Comparative analysis of optical biometers. *J Cataract Refract Surg*. 2016, 42:685-93. [10.1016/j.jcrs.2016.01.051](https://doi.org/10.1016/j.jcrs.2016.01.051)
11. Hirschschall N, Murphy S, Pimenides D, Maurino V, Findl O: Assessment of a new averaging algorithm to increase the sensitivity of axial eye length measurement with optical biometry in eyes with dense cataract. *J Cataract Refract Surg*. 2011, 37:45-9. [10.1016/j.jcrs.2010.07.023](https://doi.org/10.1016/j.jcrs.2010.07.023)
12. Sahin A, Hamrah P: Clinically relevant biometry. *Curr Opin Ophthalmol*. 2012, 23:47-53. [10.1097/ICU.0b013e32834cd63e](https://doi.org/10.1097/ICU.0b013e32834cd63e)
13. O'Gorman LD, Hogenbirk JC, Warry W: Clinical telemedicine utilization in Ontario over the Ontario Telemedicine Network. *Telemed J E Health*. 2016, 22:473-9. [10.1089/tmj.2015.0166](https://doi.org/10.1089/tmj.2015.0166)
14. Pong RW: Strategies to overcome physician shortages in northern Ontario: a study of policy implementation over 35 years. *Hum Resour Health*. 2008, 6:24. [10.1186/1478-4491-6-24](https://doi.org/10.1186/1478-4491-6-24)
15. Findl O, Kriechbaum K, Sacu S, et al.: Influence of operator experience on the performance of ultrasound biometry compared to optical biometry before cataract surgery. *J Cataract Refract Surg*. 2003, 29:1950-5.